



American Sensor Technologies, Inc.

Your sensor business partner...

June 27, 2007

To Whom it May Concern:

This letter is to confirm that AST models AST4000, 4100, 4200, 43XX, 44XX, 45XX and 4600 pressure sensors, transducers and transmitters meet the requirements of a 'Single Seal' device as defined in ANSI/ISA 12.27.01.

AST's products have been tested to UL 508 to four times the rated pressure with no rupture and are covered under UL file # E210825, E214980 and CSA file # 1393892.

Currently AST is working with UL (file # E210825) and CSA to obtain certifications for models AST4300, 43XX, 44XX, 45XX and 4600 hazardous area approved pressure sensors, transducers and transmitters.

CSA anticipates the completion of the investigation for model AST4600, by the middle of July 2007. UL testing is scheduled for completion for models AST4300, 4310, 43LP, 4400, 4410, 44LP, 4500 and 4510 by the end of August 2007.

Yours Truly,

A handwritten signature in blue ink, appearing to read 'K. Sidhu', with a horizontal line underneath.

Karmjit S. Sidhu
Vice President of Business Development
American Sensor Technologies, Inc.